

EVALUATION CERTIFICATE

No. 0200-WL-07516

Object name AE 403M / GK-Mplus
Object type Non-automatic weighing indicator
Issued by Force Certification A/S

Issued in accordance with the requirements in WELMEC Guide 8.8:2017” Guide on General and Administrative Aspects of the Voluntary System of Modular Evaluation of Measuring instruments”.

In accordance with EN 45501:2015, OIML R76:2006 and WELMEC Guide 8.8:2017.

Issued to AE Adam GmbH
Instenkamp 4
24242 Felde
Germany

Manufacturer AE Adam GmbH

In respect of Weighing indicators tested as a module for a weighing instrument.

Characteristics Suitable for non-automatic weighing instruments with the following characteristics:
Weighing range: Single-interval
Accuracy class: III or IIII
Number of VSIs: $n \leq 6000$
Minimum input voltage per VSI: $\geq 1.0 \mu V$
The essential characteristics are described in the annex.

Description and documentation The weighing indicator is described and documented in the annex to this certificate.

Remarks Summary of tests involved: see annex.

This evaluation certificate cannot be quoted in an EU type examination certificate without permission of the holder of this certificate mentioned above.

The annex comprises 7 pages.

Issued on 2019-12-06

Jens Hovgård Jensen
2019-12-06

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 119-34995.90.10 and ID no.: 0200-WL-07516

Signatory: J. Hovgård Jensen