

# EVALUATION CERTIFICATE

**No. 0200-WL-07309**

**Object name**            **260 / 290 Series**  
**Object type**            **Non-automatic weighing indicator**  
**Issued by**             **Force Certification A/S**

Issued in accordance with the requirements in WELMEC Guide 8.8:2017” Guide on General and Administrative Aspects of the Voluntary System of Modular Evaluation of Measuring instruments”.

**In accordance with** EN 45501:2015, OIML R76:2006 and WELMEC Guide 8.8:2017.

**Issued to**               **Applied Weighing International Limited**  
4-8 Southview Park, Marsack Street,  
Caversham, Reading,  
Berkshire RG4 5AF  
United Kingdom

**Manufacturer**       Applied Weighing International Limited

**In respect of**         A weighing indicator tested as a module for a weighing instrument.

**Characteristics**       Suitable for non-automatic weighing instruments with the following characteristics:  
Weighing range:                               Single-interval  
Accuracy class:                                III or IIII  
Number of VSIs:                                 $n \leq 10,000$   
Minimum input voltage per VSI:          $\geq 1.0 \mu\text{V}$   
The essential characteristics are described in the annex.

**Description and documentation**       The weighing indicator is described and documented in the annex to this certificate.

**Remarks**             Summary of tests involved: see annex.

This evaluation certificate cannot be quoted in an EU type examination certificate without permission of the holder of this certificate mentioned above.

The annex comprises 8 pages.

**Issued on**            **2019-11-29**

**Jens Hovgård Jensen**  
2019-11-29

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:  
Task no.: 119-33346.90.10 and ID no.: 0200-WL-07309

**Signatory: J. Hovgård Jensen**