

# EU Type Examination Certificate

**No. 0200-NAWI-03059**

**FX50**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**      **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**      **Avery Berkel.**  
Foundry Lane,  
Smethwick  
West Midlands  
B66 2LP  
United Kingdom

**In respect of**    Non-automatic weighing instrument designated FX50 with variants of modules of load receptors and load cells.  
Accuracy class III, single-interval, multi-range or multi-interval  
Maximum capacity, Max: From 3 kg to 30 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n = 3000$  or  $2 \times 3000$ .  
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 8 pages.

**Issued on**      **2017-11-27**  
**Valid until**    **2027-11-27**



**Signatory: J. Hovgård**

FORCE Certification references:  
Task no.: 117-34793 and ID no.: 0200-NAWI-03059