

EU Type Examination Certificate

No. 0200-NAWI-05163 Revision 1

FX110 / FX110-A

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Avery Berkel.**
Foundry Lane,
Smethwick
West Midlands, B66 2LP
United Kingdom

In respect of Non-automatic weighing instrument designated FX110 / FX110-A with variants of modules of load receptors and load cells.
Accuracy class III, single-interval
Maximum capacity, Max: 6 kg to 30 kg
Verification scale interval: $e_i = \text{Max}/n$
Maximum number of verification scale intervals: $n = 3000$.
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

Note: This certificate is a revised edition which replaces previous revisions.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 9 pages.

Jens Hovgård Jensen
2019-12-11

Issued on **2019-12-11**
Valid until **2029-01-04**

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager