

EU Type Examination Certificate

No. 0200-NAWI-04485

**BC-1803/T, BC-1803/P, BC-1803/PC, BC-1803/TI,
BC-1803/PI, BC-1803/PCI**

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Básculas y Arcas Catalunya S.A. (BACSA)**
Pol. Ind. Santiga
c/Flassaders 13-15
08130 Santa Perpetua de Mogoda
Barcelona
SPAIN

In respect of Non-automatic weighing instrument designated BC-1803/T, BC-1803/P, BC-1803/PC, BC-1803/TI, BC-1803/PI and BC-1803/PCI with variants of modules of load receptors and load cells.
Accuracy class III, single-interval, dual-range or dual-interval
Maximum capacity, Max: From 3 kg to 30 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n = 3000$.
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 14 pages.

Issued on **2018-09-06**
Valid until **2028-09-06**

Jens Hovgård Jensen
2018-09-06

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 118-29062 and ID no.: 0200-NAWI-04485

Signatory: J. Hovgård Jensen