

# EU Type Examination Certificate

**No. 0200-NAWI-6589**

**BI-1905 / BI-1905/C**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**        **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**        **BASCULAS Y ARCAS CATALUNYA S.A. (BACSA).**  
Pol. Ind santiaga. c/Flassaders 13-15, Nave 1,  
08130 Sta. Perpetua de Mogoda  
Barcelona  
SPAIN

**In respect of**    Non-automatic weighing instrument designated BI-1905 / BI-1905/C with variants of modules of load receptor and load cell.  
Accuracy class III, single-interval or multi-interval (2 intervals)  
Maximum capacity, Max: From 3 kg to 25 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n_i = \leq 3000$   
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 7 pages.

**Issued on**        **2019-05-20**  
**Valid until**     **2029-05-20**

**Jens Hovgård Jensen**  
2019-05-20

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:  
Task no.: 119-27725.90.20 and ID no.: 0200-NAWI-06589

**Signatory: J. Hovgård Jensen**