

EU Type Examination Certificate

No. 0200-NAWI-03632

BD10-SD / BS10-SD

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Birotehna doo.**
Karadjordjeva 21,
11300 Smederevo
SERBIA

In respect of Non-automatic weighing instrument designated BD10-SD / BS10-SD with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III
Maximum capacity, Max: From 1 kg up to 150 000 kg
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n \leq 3000$
(however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015, OIML R76:2006 and WELMEC 2.1:2001.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 19 pages.

Issued on **2018-02-07**
Valid until **2028-02-07**



Signatory: J. Hovgård

FORCE Certification references:
Task no.: 117-22038 and ID no.: 0200-NAWI-03632