



We help ideas meet the real world

# EC Type-Approval Certificate

## No. DK 0199.377 Revision 1

### MWI / MCI / LKI / EWI

### NON-AUTOMATIC WEIGHING INSTRUMENT

**Issued by** DELTA Danish Electronics, Light & Acoustics  
EU - Notified Body No. 0199

In accordance with the requirements for the non-automatic weighing instrument of EC Council Directive 2009/23/EC.

**Issued to** **Bosche GmbH & Co. KG**  
Reselager Rieden 3  
49401 Damme  
Germany

**In respect of** Non-automatic weighing instrument designated MWI / MCI / LKI / EWI all based on BDU10-M / BDU10-S analog-to-digital data processing unit with variants of modules of load receptors, load cells and peripheral equipment.  
Accuracy class III and IIII  
Maximum capacity, Max: From 1 kg up to 999 950 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n_i \leq 10000$  for single-interval and multi-range (however, dependent on environment and the composition of the modules).  
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015, OIML R76:2006 and WELMEC Guide 2.1:2001

**Note: This certificate is a revised edition, which replaces previous revisions of the certificate**

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 18 pages.

**Issued on** 2015-11-30  
**Valid until** 2023-06-17

  
**Signatory: J. Hovgård**

**DELTA**  
Venlighedsvej 4  
2970 Hørsholm  
Denmark  
  
Tel. (+45) 72 19 40 00  
Fax (+45) 72 19 40 01  
www.delta.dk  
VAT No. DK 12275110