

EU Type-Examination Certificate

No. 0200-NAWI-03187 Revision 1

185

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Cardinal Scale Manufacturing Company**
203 East Daugherty Street
Webb City
Missouri 64870
USA

In respect of Non-automatic weighing instrument designated 185 with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III, single-interval
Maximum capacity, Max: $n \times e$
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n \leq 6000$
(however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of EN 45501:2015, OIML R76:2006. and WELMEC 2.1:2001.

The principal characteristics and examination conditions are set out in the descriptive annex to this certificate.

The annex comprises 13 pages.

Issued on **2018-12-03**
Valid until **2028-10-25**

Jens Hovgård Jensen
2018-12-03

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 118-35298.90.10 and ID no.: 0200-NAWI-05272

Signatory: J. Hovgård