

EVALUATION CERTIFICATE

No. 0200-WL-08648

Object name	SmartCAN
Object type	Non-automatic weighing transmitter
Issued by	FORCE Certification EU - Notified Body No. 0200
In accordance with	EN 45501:2015, OIML R76:2006 and WELMEC Guide 8.8:2017 on metrological aspects of non-automatic weighing instruments.
Fractional factor (p_i)	0.5 (refer to 3.10.2.1 of EN 45501:2015).
Issued to	Cardinal Scale Manufacturing Company 102 East Daugherty Street, Webb City, Missouri 64870, USA
Manufacturer	Cardinal Scale Manufacturing Company.
In respect of	Non-automatic weighing transmitter.
Characteristics	Suitable for a non-automatic weighing instruments with the following characteristics: Weighing range: Single-interval. Accuracy class: III or IIII Verification scale interval: $e_i =$ Max_i/n_i Maximum number of verification scale intervals: $n_i =$ ≤ 6000 (Class III) / ≤ 1000 (Class IIII) Minimum input voltage per VSI: $0.5 \mu V$ The essential characteristics are described in the annex.
Description and documentation	The non-automatic weighing transmitter is described and documented in the annex to this certificate.
Remarks	Summary of tests involved: See the annex to this certificate.

This evaluation certificate cannot be quoted in an EU type examination certificate without permission from the holder of the certificate mentioned above.

The annex comprises 9 pages.

Issued on **2020-07-03**

Jens Hovgård Jensen
2020-07-03

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 119-35654.90.10 and ID no.: 0200-WL-08648

Signatory: Jens Hovgård Jensen