

# EU Type Examination Certificate

**No. 0200-NAWI-06532**

**CT100 Series**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**        **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**        CAS Corporation  
#262, Geurugogae-ro, Gwangjeok-myeon,  
Yangju-si, Gyeonggi-do  
REPUBLIC OF KOREA

**In respect of**    Non-automatic weighing instrument designated CT100 with variants of modules of load receptors and load cells.  
Accuracy class III, single-interval or multi-interval  
Maximum capacity, Max: 6 kg to 30 kg  
Verification scale interval:  $e \geq 1$  g  
Verification scale intervals:  $n \leq 3000$  or  $n_i \leq 2 \times 3000$   
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 12 pages.

**Issued on**        **2019-10-31**  
**Valid until**     **2029-10-31**

**Jens Hovgård Jensen**  
2019-10-31

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:  
Task no.: 119-23195.90.20 and ID no.: 0200-NAWI-06532

**Signatory: J. Hovgård Jensen**