

# EVALUATION CERTIFICATE

**No. 0200-WL-07308**

**Object name**                **EBI / DBI**  
**Object type**                **Weighing indicator for a non-automatic weighing instrument**  
**Issued by**                **Force Certification A/S**

Issued in accordance with the requirements in WELMEC Guide 8.8:2017 "Guide on General and Administrative Aspects of the Voluntary System of Modular Evaluation of Measuring instruments".

**In accordance with**    OIML R76:2006 and EN 45501:2015 on metrological aspects of non-automatic weighing instruments.

**Issued to**                **CAS Corporation**  
#262, Geurugogae-ro, Gwangjeok-myeon,  
Yangju-si, Gyeonggi-do  
REPUBLIC OF KOREA

**Manufacturer**        **CAS Corporation**

**In respect of**        Non-automatic weighing indicator

**Characteristics**        The EBI / DBI indicators have the following characteristics:  
Accuracy class:                                III or IIII  
Weighing range:                                Single-interval or multi-interval  
Number of VSIs:                                 $n \leq 3,000$   
Minimum input voltage per VSI:         $1.2 \mu V$   
The essential characteristics are described in the annex.

**Description and documentation**        The weighing indicators are described and documented in the annex to this this certificate.

**Remarks**                The conformity was established by the reports listed in the annex.

This evaluation certificate cannot be quoted in an EU type examination certificate without permission of the holder of this certificate mentioned above.

The annex comprises 8 pages.

**Issued on**                **2019-10-31**

**Jens Hovgård Jensen**  
2019-10-31

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager