

EU Type Examination Certificate

No. 0200-NAWI-06588

PS50-M / PS50-M CW

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **DIBAL S.A.**
c/Astintze, 26,
Poligono Industrial Neinver
48160 Derio
SPAIN

In respect of Non-automatic weighing instrument designated PS50-M / PS50-M CW with variants of modules of load receptor and load cell.
Accuracy class III, single-interval or multi-interval (2 intervals)
Maximum capacity, Max: From 3 kg to 25 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n_i = \leq 3000$
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 7 pages.

Issued on **2019-05-20**
Valid until **2029-05-20**

Jens Hovgård Jensen
2019-05-20

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 119-2775.90.10 and ID no.: 0200-NAWI-06588

Signatory: J. Hovgård Jensen