



We help ideas meet the real world

EC Type Examination Certificate

DK0199.399

BOS

AUTOMATIC GRAVIMETRIC FILLING INSTRUMENT

Issued by DELTA Danish Electronics, Light & Acoustics
EU - Notified Body No. 0199

In accordance with the requirements for the automatic weighing instrument of Directive 2004/22/EC of the European Parliament and Council on Measuring Instruments (MID).

Issued to Elektroniczne Wagi Przemyslowe
Ul. Zacna 31
80-283 Gdansk
Poland

In respect of An automatic gravimetric filling instrument designated BOS with variants of modules of load receptors, load cells and peripheral equipment.
Reference accuracy class, Ref(x): 0.2
Accuracy class: 0.2, 0.5, 1 or 2.
Maximum capacity: $6 \text{ kg} \leq \text{Max} \leq 1500 \text{ kg}$.
Verification scale interval: $2 \text{ g} \leq d \leq 500 \text{ g}$
Maximum number of verification scale intervals: $n \leq 3000$ (however dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in Annex 1 and the specific requirements in Annex MI-006, chapter I & III of the Directive 2004/22/EC is met by the application of OIML R61-1:2004, OIML D11:2004 section 12 & 13 with severity level 3, WELMEC Guide 7.2:2009, and WELMEC Guide 8.16-3:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 13 pages.

Issued on 2013-10-03
Valid until 2023-10-03


Signatory: J. Hovgård

DELTA
Venlighedsvej 4
2970 Hørsholm
Denmark

Tel. +45 72 19 40 00
Fax +45 72 19 40 01
www.delta.dk
VAT No. DK 12275110