



EU Type Examination Certificate

DK0199.636 Revision 1

BOS

AUTOMATIC GRAVIMETRIC FILLING INSTRUMENT

Issued by DELTA
EU - Notified Body No. 0199

In accordance with the requirements for the automatic weighing instruments in Directive 2014/32/EU of the European Parliament and Council of February 26, 2014 on Measuring Instruments (MID).

Issued to Elektroniczne Wagi Przemyslowe
Ul. Zacna 31
80-283 Gdansk
Poland

In respect of An automatic gravimetric filling instrument designated BOS with variants of modules of load receptors, load cells and peripheral equipment.
Reference accuracy class, Ref(x): 0.1
Accuracy class: 0.2, 0.5, 1 or 2.
Maximum capacity: $0.5 \text{ kg} \leq \text{Max} \leq 50000 \text{ kg}$.
Verification scale interval: $0.5 \text{ g} \leq d \leq 20 \text{ kg}$
Maximum number of verification scale intervals: $n \leq 3000$ (however dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in Annex 1 and the specific requirements in Annex VIII (MI-006), chapter I & III of the Directive 2014/32/EU is met by the application of OIML R61-1:2004 and WELMEC Guide 7.2.

Note: This revised edition replaces earlier revisions of the certificate.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 12 pages.

Issued on 2017-02-01
Valid until 2027-01-24


Signatory: J. Hovgård

DELTA
Venlighedsvej 4
2970 Hørsholm
Denmark

Tel. +45 72 19 40 00
Fax +45 72 19 40 01
www.delta.dk
VAT No. DK55117314