



We help ideas meet the real world

EVALUATION CERTIFICATE

No. DK0199-16.05

Instrument type LDU 78.1 version 3

Test item device Analog data processing device

Issued by DELTA
EU Notified Body No. 0199

In accordance with EN 45501:2015, OIML R76:2006 and WELMEC Guide 8.8:2011 on metrological aspects of non-automatic weighing instruments.

Fractional factor (p_i) 0.5 (refer to 3.10.2.1 of EN 45501:2015).

Issued to Hauch & Bach ApS
Femstykket 6
3540 Lyngø
Denmark

Manufacturer Hauch & Bach ApS

In respect of An analog data processing device tested as a module for a weighing instrument.

Characteristics Suitable as a non-automatic weighing instruments with the following characteristics:
Weighing range: Single-interval or multi-range or multi-interval
Accuracy class: III or IIII
Verification scale interval: $e =$ Max/n
Maximum number of verification scale intervals: $n =$ 10,000
Minimum input voltage per VSI: 0.3 μ V
The essential characteristics are described in the annex.

Description and documentation The analog data processing device is described and documented in the annex to this certificate.

Remarks Summary of tests involved: See test report no. DANAK-1911008.

This test certificate cannot be quoted in an EU type approval certificate without permission from the holder of the certificate mentioned above.

The annex comprises 6 pages.

Issued on 2016-07-19


Signatory: J. Hovgård

DELTA
Venlighedsvej 4
2970 Hørsholm
Denmark
Tel. (+45) 72 19 40 00
Fax (+45) 72 19 40 01
www.delta.dk
VAT No. DK 12275110