

# EU Type Examination Certificate

**No. 0200-NAWI-08454**

**Micro S29 / SW / SW-II / SW-III**  
**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**        **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**        **Joaquim Gabriel De Barros Rodrigues Unipessoal LDA**  
Rua Do Ermo 523 Grijo  
4415 460 Vila Nova De Gaia, Porto  
Portugal

**In respect of**    Non-automatic weighing instrument designated Micro S29 / SW / SW-II / SW-III with variants of modules of load receptors and load cells.  
Accuracy class III, single-interval, dual-range or dual-interval  
Maximum capacity, Max: From 1.5 kg up to 30 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n_i \leq 3000$ .  
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 11 pages.

**Issued on**        **2020-05-18**  
**Valid until**     **2030-05-18**

**Jens Hovgård Jensen**  
2020-05-18

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:  
Task no.: 120-26806.90.20 and ID no.: 0200-NAWI-08454

**Signatory: Jens Hovgård Jensen**