

EVALUATION CERTIFICATE

No. 0200-WL-04954

Instrument type **W100 / W200 / W200IP64 / W200IP67 / WDOS /
WDESK / WDESKLIGHT / WINOX / WTAB /
W200BOX / W200BOXEC / ADPEW200 / WLIGHT**

Test item device Non-automatic weighing indicator

Issued by **FORCE Certification**
EU - Notified Body No. 0200

Issued in accordance with the requirements in WELMEC Guide 8.8:2017" Guide on General and Administrative Aspects of the Voluntary System of Modular Evaluation of Measuring instruments".

In accordance with EN 45501:2015 and OIML R76:2006.

Fractional factor (p_i) 0.5 (refer to 3.10.2.1 of EN 45501:2015).

Issued to **Laumas Elettronica SRL.**
Via Primo Maggio 6
I - 43022 Montechiarugolo (PR)
ITALY

Manufacturer Laumas Elettronica SRL.

In respect of A family of indicators tested as a module of a non-automatic weighing instrument.

Characteristics Suitable as a non-automatic weighing instrument with the following characteristics:
Self indicating with single-interval, multi-interval or multi-range
Accuracy class III or IIII
Verification scale interval: $e_i = \text{Max}_i/n_i$
Maximum number of
verification scale intervals: $n = 10000$ for class III
 $n = 1000$ for Class IIII
Min. input voltage per VSI: $0.2 \mu\text{V}$
The essential characteristics are described in the annex.

Description and documentation The indicator is described and documented in the annex to this certificate.

Note: This certificate is a revision of test certificate DK0199-R76-11.01 rev.1.

This evaluation certificate cannot be quoted in an EU type approval certificate without permission from the holder of the certificate mentioned above.

The annex comprises 7 pages.

Issued on 2019-01-10

FORCE Certification references:

Task no.: 118-30564.90.20 and ID no.: 0200-WL-04954

Jens Hovgård Jensen
2019-01-10

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

Signatory: J. Hovgård Jensen