



We help ideas meet the real world

EC Type-Approval certificate

No. DK 0199.31 Revision 5

M1100-Cx

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by DELTA Danish Electronics, Light & Acoustics
EU - Notified Body No. 0199

In accordance with the requirements for the non-automatic weighing instrument of EC Council Directive 2009/23/EC.

Issued to Marel ehf
Austurhraun 9
210 Gardabaer
Iceland

In respect of Non-automatic weighing instrument designated M1100-Cx with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III or Class III
Maximum capacity, Max: From 0.3 kg to 375,000 kg
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n = 7,500$ for class III, 1,000 (class III), (however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:1992.

Note: This certificate is a revised edition, which replaces previous revisions and extends the validation period of the certificate.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 21 pages.

Issued on 2016-02-08
Valid until 2026-02-08


Signatory: J. Hovgård

DELTA
Venlighedsvej 4
2970 Hørsholm
Denmark

Tel. +45 72 19 40 00
Fax +45 72 19 40 01
www.delta.dk
VAT No. DK 12275110