

EU Type Examination Certificate

No. 0200-NAWI-05309

T28

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **MURBAY ITHALAT IHRACAT ELEKTRONIK SAN. TIC.LTD.STI.**
Kircaali Mah. Celal Bayar Cad. No: 137/A Osmangazi,
Bursa
TURKEY

In respect of Non-automatic weighing instrument designated T28 with variants of modules of load receptor and load cell.
Accuracy class III, single-interval or multi-interval (2 intervals)
Maximum capacity, Max: From 3 kg to 30 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n_i = \leq 3000$
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 7 pages.

Issued on **2019-02-01**
Valid until **2029-02-01**

Jens Hovgård Jensen
2019-02-01

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 118-34735.90.25 and ID no.: 0200-NAWI-05309

Signatory: J. Hovgård Jensen