

EU Type Examination Certificate

No. 0200-NAWI-05310

S29

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **MURBAY ITHALAT IHRACAT ELEKTRONIK SAN. TIC.LTD.STI.**
Kircaali Mah. Celal Bayar Cad. No: 137/A Osmangazi,
Bursa
TURKEY

In respect of Non-automatic weighing instrument designated S29 with variants of modules of load receptors and load cells.
Accuracy class III, single-interval, dual-range or dual-interval
Maximum capacity, Max: From 1.5 kg up to 30 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n_i \leq 3000$.
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 9 pages.

Issued on **2019-02-01**
Valid until **2019-02-01**

Jens Hovgård Jensen
2019-02-01

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 118-34735.90.30 and ID no.: 0200-NAWI-05310 **Signatory: J. Hovgård Jensen**