

# EU Type Examination Certificate

**No. 0200-NAWI-04255**

**WS**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**        **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**        **PMC Milliot**  
Allée des Epinettes,  
Bat. N° 9 Z.I.Nord  
77200 Torcy  
FRANCE

**In respect of**    Non-automatic weighing instrument designated WS with variants of modules of load receptors and load cells.  
Accuracy class III, single-interval, dual-range or dual-interval  
Maximum capacity, Max: From 1.5 kg up to 30 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n_i \leq 3000$ .  
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 10 pages.

**Issued on**        **2018-06-11**  
**Valid until**     **2028-06-11**

  
**Signatory: J. Hovgård**

FORCE Certification references:  
Task no.: 118-25907 and ID no.: 0200-NAWI-04255