

EU Type Examination Certificate

No. 0200-NAWI-04256

TNP / TSP / ZNW

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **PMC Milliot**
Allée des Epinettes,
Bat. N° 9 Z.I.Nord
77200 Torcy
FRANCE

In respect of Non-automatic weighing instrument TNP / TSP / ZNW with variants of modules of load receptors and load cells.
Accuracy class III, single-interval, dual-range or dual-interval
Maximum capacity, Max: From 3 kg to 30 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n = 3000$ or 2×3000 .
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

Note: This certificate is a revised edition of DK0199.641.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 11 pages.

Issued on **2018-06-11**
Valid until **2027-03-08**


Signatory: J. Hovgård

FORCE Certification references:
Task no.: 118-25907 and ID no.: 0200-NAWI-04256