

# EU Type Examination Certificate

**No. 0200-NAWI-02963**

**PT610 / PT620 / PT630**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**        **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**        **PT Limited**  
5B Marken Place, Glenfield 0627,  
Auckland  
New Zealand

**In respect of**    Non-automatic weighing instrument designated PT610 / PT620 / PT630 with variants of modules of load receptors, load cells and peripheral equipment.  
Accuracy class: III and IIII  
Maximum capacity, Max: From 1 kg up to 300 000 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n_i \leq 10000$  for single-interval / multi-range / multi-interval (however, dependent on environment and the composition of the modules).  
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and conditions for certification are set out in the descriptive annex to this certificate.

The annex comprises 17 pages.

**Issued on**        **2017-09-22**  
**Valid until**     **2027-09-22**



**Signatory: J. Hovgård**

FORCE Certification references:  
Task no.: 117-25861.35 and ID no.: 0200-NAWI-02963