

EU Type Examination Certificate

No. 0200-NAWI-04386

PT252 / PT253 / PT272

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **PT Limited**
5B Marken Place, Glenfield 0627,
Auckland
New Zealand

In respect of Non-automatic weighing instrument designated PT252 / PT253 / PT272 with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III and IIII
Maximum capacity, Max: From 1 kg up to 300 000 kg
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n \leq 6000$ for single-interval and multi-range (however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

Note: This certificate is a revised edition of DK0199.559 rev.1

The principal characteristics and conditions for certification are set out in the descriptive annex to this certificate.

The annex comprises 18 pages.

Issued on **2018-06-07**
Valid until **2025-11-03**


Signatory: J. Hovgård

FORCE Certification references:
Task no.: 118-23422.25 and ID no.: 0200-NAWI-04386