

EU Type Examination Certificate

No. 0200-NAWI-04523

PM5 / PM6

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Tehno mgm d.o.o.**
Vojvode Petka 2/9
Belgrade
Serbia

In respect of Non-automatic weighing instrument designated PM5 / PM6 with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III
Maximum capacity, Max: From 60 kg to 300 kg
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n = 3000$, single-interval (however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of EN 45501:2015 and OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 13 pages.

Issued on **2018-07-05**
Valid until **2028-07-05**


Signatory: J. Hovgård