

# EVALUATION CERTIFICATE

**No. 0200-WL-10218**

<b>Object name</b>	<b>I400E / I750E / I510 / I700 / AS</b>	
<b>Object type</b>	<b>Non-automatic weighing indicator</b>	
<b>Issued by</b>	<b>FORCE Certification</b> EU - Notified Body No. 0200	
<b>In accordance with</b>	EN 45501:2015, OIML R76:2006 and WELMEC Guide 8.8:2017 on metrological aspects of non-automatic weighing instruments.	
<b>Fractional factor (p<sub>i</sub>)</b>	0.5 (refer to 3.10.2.1 of EN 45501:2015).	
<b>Issued to</b>	<b>Ascell Sensor S.L.</b> Congost Av. 56 – Nave 3 08760 Martorell – Barcelona Spain	
<b>Manufacturer</b>	Dini Argeo S.r.l.	
<b>In respect of</b>	Non-automatic weighing indicators.	
<b>Characteristics</b>	Suitable as a non-automatic weighing instruments with the following characteristics: Weighing range: Single-interval or multi-interval or multi-range. Accuracy class: III or IIII Verification scale interval: $e_i =$ $Max_i/n_i$ Maximum number of verification scale intervals: $n_i =$ 10,000 per interval/range Minimum input voltage per VSI: 0.3 $\mu$ V The essential characteristics are described in the annex.	
<b>Description and documentation</b>	The non-automatic weighing indicators are described and documented in the annex to this certificate.	
<b>Remarks</b>	Summary of tests involved: See the annex to this certificate.	

This evaluation certificate cannot be quoted in an EU type examination certificate without permission from the holder of the certificate mentioned above.

The annex comprises 57 pages.

**Issued on** 2021-02-22

**Jens Hovgård Jensen**  
2021-02-22

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:

Task no.: 121-22544.90.10 and ID no.: 0200-WL-10218

**Signatory: Jens Hovgård Jensen**