

EU Type Examination Certificate

No. 0200-NAWI-16327

X-Pro Series

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Avery Berkel**
Foundry Lane,
Smethwick,
West Midlands B66 2LP
UNITED KINGDOM

In respect of Non-automatic weighing instrument family designated X-Pro Series with variants of modules of load receptors and load cells.
Accuracy class III, single-interval or multi-interval
Maximum capacity, Max: From 6 kg to 30 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n = 6000$ or 2×3000 .
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 14 pages.

Issued on **2024-01-16**
Valid until **2034-01-16**

Jens Hovgård Jensen
2024-01-16

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 123-32880.90.10 and ID no.: 0200-NAWI-16327-1

Signatory: Jens Hovgård Jensen