



We help ideas meet the real world

EC Type-Approval Certificate

No. DK 0199.417 Revision 1

PD300M

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by DELTA Danish Electronics, Light & Acoustics
EU - Notified Body No. 0199

In accordance with the requirements for the non-automatic weighing instrument of EC Council Directive 2009/23/EC.

Issued to Cardinal Scale Manufacturing Company
203 East Daugherty
P.O. Box 151
Webb City, MO 64870
USA

In respect of Non-automatic weighing instrument designated PD300M.
Accuracy class III, single-interval
Maximum capacity, Max: 250 kg
Verification scale interval: $e \geq 0.1$ kg
Maximum number of verification scale intervals: $n = 2500$ (however, dependent on environment).

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:1992/AC:1993 and OIML R76:2006.

Note: This certificate is a revised edition which replaces previous revisions.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 6 pages.

Issued on 2015-04-16
Valid until 2023-08-26


Signatory: J. Hovgård

DELTA
Venlighedsvej 4
2970 Hørsholm
Denmark
Tel. (+45) 72 19 40 00
Fax (+45) 72 19 40 01
www.delta.dk
VAT No. DK 12275110