

EU Type Examination Certificate

No. 0200-NAWI-08685

WPB

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Cardinal Scale Manufacturing Company**
102 East Daugherty Street,
Webb City, Missouri 64870
USA

In respect of Non-automatic weighing instrument designated with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III, single-interval or dual range
Maximum capacity, Max: From 3 kg up to 30 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n_i = 3000$.
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 8 pages.

Issued on **2020-07-03**
Valid until **2030-07-03**

Jens Hovgård Jensen
2020-07-03

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 120-28285.90.10 and ID no.: 0200-NAWI-08685

Signatory: Jens Hovgård Jensen