

# EU Type Examination Certificate

**No. 0200-NAWI-04801**

**225**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**       **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**       **Cardinal Scale Manufacturing Company**  
102 East Daugherty  
P.O. Box 151  
Webb City, MO 64870  
USA

**In respect of**   Non-automatic weighing instrument designated 225 with variants of modules of load receptors, load cells and peripheral equipment.  
Accuracy class III or IIII, single-interval or dual-interval  
Maximum capacity, Max: From 1 kg up to 999 999 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n_i = 10000$  for Class III,  
 $n_i = 1000$  for class IIII (however, dependent on environment and the composition of the modules).  
Variants of modules and conditions for the composition of modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

**Note: This certificate is a revised edition extending the validation period.**

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 13 pages.

**Issued on**       **2024-02-25**  
**Valid until**   **2034-02-25**

**Jens Hovgård Jensen**  
2024-02-19

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:

Task no.: 124-23035.90.10 and ID no.: 0200-NAWI-04801-2

**Signatory: Jens Hovgård Jensen**