

EU Type Examination Certificate

No. 0200-NAWI-05287 Rev. 1

XE / XE PLUS

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to CAS Corporation
#262, Geurugogae-ro, Gwangjeok-myeon,
Yangju-si, Gyeonggi-do
REPUBLIC OF KOREA

In respect of Non-automatic weighing instrument designated XE / XE PLUS with variants of modules of load receptors and load cells.
Accuracy class II, single-interval
Maximum capacity, Max: 300 g to 6000 g
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals:
XE: $n = 6000$ or $n = 7500$
XE PLUS: $n = 30000$
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

Note: This certificate is a revised edition which replaces previous revisions.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 12 pages.

Issued on **2019-05-23**
Valid until **2025-04-13**

Jens Hovgård Jensen
2019-05-23

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 118-34761.90.10 and ID no.: 0200-NAWI-06054 **Signatory: J. Hovgård Jensen**