

EU Type Examination Certificate

No. 0200-NAWI-12918

CL5500N Series

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to CAS Corporation
#262, Geurugogae-ro, Gwangjeok-myeon,
Yangju-si, Gyeonggi-do
REPUBLIC OF KOREA

In respect of Non-automatic weighing instrument designated CL5500N with variants of modules of load receptors and load cells.
Accuracy class III, single-interval or multi-interval
Maximum capacity, Max: 6 kg to 30 kg
Verification scale interval: $e \geq 1$ g
Verification scale intervals: $n \leq 3000$ or $n_i \leq 2 \times 3000$
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 12 pages.

Issued on **2022-06-16**
Valid until **2032-06-16**

Jens Hovgård Jensen
2022-06-16

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 121-31870.90.10 and ID no.: 0200-NAWI-12918

Signatory: J. Hovgård Jensen