

# EVALUATION CERTIFICATE

## No. 0200-WL-07308 Revision 1

**Object name** EBI / DBI  
**Object type** Weighing indicator for a non-automatic weighing instrument  
**Issued by** Force Certification A/S

Issued in accordance with the requirements in WELMEC Guide 8.8:2017 "Guide on General and Administrative Aspects of the Voluntary System of Modular Evaluation of Measuring instruments".

**In accordance with** OIML R76:2006 and EN 45501:2015 on metrological aspects of non-automatic weighing instruments.

**Issued to** **CAS Corporation**  
#262, Geurugogae-ro, Gwangjeok-myeon,  
Yangju-si, Gyeonggi-do  
REPUBLIC OF KOREA

**Manufacturer** **CAS Corporation**

**In respect of** Non-automatic weighing indicator

**Characteristics** The EBI / DBI indicators have the following characteristics:  
Accuracy class: III or IIII  
Weighing range: Single-interval or multi-interval  
Number of VSIs:  $n \leq 3,000$   
Minimum input voltage per VSI:  $1.2 \mu\text{V}$   
The essential characteristics are described in the annex.

**Description and documentation** The weighing indicators are described and documented in the annex to this this certificate.

**Remarks** The conformity was established by the reports listed in the annex.

**Note: This certificate is a revised edition which replaces previous revisions.**

This evaluation certificate cannot be quoted in an EU type examination certificate without permission of the holder of this certificate mentioned above.

The annex comprises 8 pages.

**Issued on** 2021-08-13

**Jens Hovgård Jensen**  
2021-08-13

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:

Task no.: 121-27661.90.40 and ID no.: 0200-WL-07308-1

**Signatory: Jens Hovgård Jensen**