

EU Type Examination Certificate

No. 0200-NAWI-02943 Revision 1

EW100 / EW200

NON-AUTOMATIC WEIGHING INDICATOR

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Chengdu Pris Electronic Co., Ltd.**
C3-2F, Mould Tooling Industry Park,
West Hi-Tech. Zone,
Chengdu,
China

In respect of Non-automatic weighing instrument designated EW100 / EW200 with variants of modules of load receptors and load cells.
Accuracy class III, multi-range (2 ranges)
Maximum capacity, Max: From 3 kg up to 30 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n_i \leq 3000$.
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

Note: This certificate is a revised edition extending the validation period.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 9 pages.

Issued on **2022-12-06**
Valid until **2032-12-10**

Jens Hovgård Jensen
2022-12-06

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 122-25557.90.10 and ID no.: 0200-NAWI-02943-2

Signatory: Jens Hovgård Jensen