

# EU Type Examination Certificate

**No. 0200-NAWI-11522**

**EP100 / EP110 / EP200 / EP210**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**        **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**        **Chengdu Pris Electronic Co., Ltd.**  
C3-2F, Mould Tooling Industry Park,  
West Hi-Tech. Zone,  
Chengdu,  
China

**In respect of**    Non-automatic price computing weighing instrument designated EP100 / EP110 / EP200 / EP210 with variants of modules of load receptors and load cells.  
Accuracy class III, single-interval, multi-interval (2 intervals) or multi-range (2 ranges)  
Maximum capacity, Max: From 3 kg up to 30 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n_i \leq 3000$ .  
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

**Note: This certificate is a revised edition of certificate DK0199.396 revision 2**

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 12 pages.

**Issued on**        **2021-09-30**  
**Valid until**     **2023-09-30**

**Jens Hovgård Jensen**  
2021-09-30

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:

Task no.: 121-31466.90.10 and ID no.: 0200-NAWI-11522-1

**Signatory: Jens Hovgård Jensen**