DANAK
PROD Reg. nr 7026

FORCE Certification

EU Type Examination Certificate

No. 0200-NAWI-13929 Revision 2

IW100A / IW300A / IW500A / IW110A / IW310A / IW510A

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Chengdu Pris Electronic Co., Ltd.,**
Building No.5,No.20 of Xixin Avenue,
West Hi-Tech Zone,
Chengdu
China

In respect of Non-automatic weighing instrument designated IW100A / IW300A / IW500A / IW110A / IW310A / IW510A with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III
Maximum capacity, Max: ≤ 375 000 kg
Verification scale interval: e = Max / n
Maximum number of verification scale intervals:
Models IW100A / IW300A / IW500A: n ≤ 7500
Models IW110A / IW310A / IW510A: n ≤ 3000
(however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015, OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 14 pages.

Issued on **2026-04-08**
Valid until **2033-01-30**

Michael Lang Sørensen
2026-04-08
Digitally signed by Michael Lang Sørensen
mls@forcetechnology.com
Project Manager

FORCE Certification references:
Task no.: 124-26314.90.10 and ID no.: 0200-NAWI-21754-2 **Signatory: Michael Lang Sørensen**

This EU type examination certificate may not be reproduced except in full, without written permission by FORCE Certification A/S