

EU Type Examination Certificate

No. 0200-NAWI-13929 Revision 1

IW100A / IW300A / IW500A

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Chengdu Pris Electronic Co., Ltd.,**
C3-2F, Mould Tooling Industry Park,
West Hi-Tech Zone,
Chengdu
China

In respect of Non-automatic weighing instrument designated IW100A / IW300A / IW500A with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III
Maximum capacity, Max: $\leq 375\ 000$ kg
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n \leq 7500$ (however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015, OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 14 pages.

Issued on **2023-03-02**
Valid until **2033-01-30**

Jens Hovgård Jensen
2023-03-02

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 121-27384.90.10 and ID no.: 0200-NAWI-13929-2

Signatory: Jens Hovgård Jensen