DANAK
PROD Reg. nr 7026

FORCE Certification

EU Type Examination Certificate

No. 0200-NAWI-18152

HT-xxxNA / HT-xxxSA

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to Chengdu Pris Electronic Co., Ltd.
C3-2F, Mould Tooling Industry Park
West Hi-Tech Zone
Chengdu
CHINA

In respect of Non-automatic weighing instrument designated HT-xxxNA / HT-xxxSA with variants of modules of load receptors and load cells.
Accuracy class II, single-interval
Maximum capacity, Max: 300 g to 6000 g
Verification scale interval: e = Max/ n
Maximum number of verification scale intervals: n = 6000 or n = 7500
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

Note: This certificate is a revised edition, which replaces DK0199.458 revision 2 and extend the validation period.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 13 pages.

Issued on **2024-10-17**
Valid until **2034-10-17**

Jens Hovgård Jensen
2024-10-15

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 124-32079.90.10 and ID no.: 0200-NAWI-18152-1

Signatory: Jens Hovgård Jensen