

# EU Type Examination Certificate

**No. 0200-NAWI-05869 Revision 1**

**ZTP / ZSP / ZHW-2 / ZHW-7 / ZHW-20 / ZHC / WTP / WSP /  
UTP / USP / UHP / UHPS / UTW-2 / UTW-7 / UTW-20 / UTC  
NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**       **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**       **CUSTOM SPA**  
Via Berettine 2/B,  
43010 Fontevivo (PR)  
Italy

**In respect of** Non-automatic weighing instrument designated ZTP / ZSP / ZHW-2 / ZHW-7 / ZHW-20 / ZHC / WTP / WSP / UTP / USP / UHP / UHPS / UTW-2 / UTW-7 / UTW-20 / UTC with variants of modules of load receptors and load cells.  
Accuracy class III, single-interval, dual-range or dual-interval  
Maximum capacity, Max: From 3 kg to 30 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n = 3000$  or  $2 \times 3000$ .  
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

**Note: This certificate is a revised edition which replaces previous revisions.**

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 18 pages.

**Issued on**       **2021-11-23**  
**Valid until**   **2029-03-28**

**Jens Hovgård Jensen**  
2021-11-23

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:  
Task no.: 121-33719.90.30 and ID no.: 0200-NAWI-05869-2

**Signatory: Jens Hovgård Jensen**