DANAK
PROD Reg. nr 7026

FORCE Certification

EU Type Examination Certificate

No. 0200-NAWI-17168 Revision 1

ST510P

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Changzhou LINWT Instrument Co., Ltd.**
No.2 Building 56, Tusincere Science Park
No. 8 beitanghe Road, Tianning District,
Changzhou City
China

In respect of Non-automatic weighing instrument designated ST510P with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class III and IIII
Single interval
Maximum capacity, Max: ≤ 900 000 kg
Verification scale interval: e = Max / n
Maximum number of verification scale intervals: n ≤ 6000 (however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015, OIML R76:2006
.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 11 pages.

Issued on **2024-10-18**
Valid until **2034-05-30**

Jens Hovgård Jensen
2024-10-18

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:
Task no.: 124-30259.90.10 and ID no.: 0200-NAWI-18220-1

Signatory: Jens Hovgård Jensen