



We help ideas meet the real world

EC Type-Approval Certificate

No. DK 0199.474 Revision 1

HT-xxxNA / HT-xxxSA

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by DELTA Danish Electronics, Light & Acoustics
EU - Notified Body No. 0199

In accordance with the requirements for the non-automatic weighing instrument of EC Council Directive 2009/23/EC.

Issued to Dikomsan Elektronik Sanayi ve Ticaret Ltd. Sti.
Oto Sanayi Sitesi, Menderes Cad. No:19
4.Levent / Istanbul
TURKEY

In respect of Non-automatic weighing instrument designated HT-xxxNA / HT-xxxSA with variants of modules of load receptors and load cells.
Accuracy class II, single-interval
Maximum capacity, Max: 300 g to 6000 g
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n = 6000$ or $n = 7500$
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:1992/AC:1993 and OIML R76:2006

Note: This certificate is a revised edition which replaces previous revisions.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 10 pages.

Issued on 2014-12-15
Valid until 2024-08-24


Signatory: J. Hovgård

DELTA
Venlighedsvej 4
2970 Hørsholm
Denmark

Tel. (+45) 72 19 40 00
Fax (+45) 72 19 40 01
www.delta.dk
VAT No. DK 12275110