

EU Type Examination Certificate

No. 0200-NAWI-12899

HT-xxxNA / HT-xxxSA

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Dikomsan Elektronik Sanayi ve Ticaret Ltd. Sti.**
Huzur Mah. Menderes Cad. No:19 4. Levent,
Sarıyer, Istanbul
Turkey

In respect of Non-automatic weighing instrument designated HT-xxxNA / HT-xxxSA with variants of modules of load receptors and load cells.
Accuracy class II, single-interval
Maximum capacity, Max: 300 g to 6000 g
Verification scale interval: $e = \text{Max} / n$
Maximum number of verification scale intervals: $n = 6000$ or $n = 7500$
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015, OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 12 pages.

Issued on **2022-06-23**
Valid until **2032-06-23**

Jens Hovgård Jensen
2022-06-23

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 122-25555.90.10 and ID no.: 0200-NAWI-12899-1

Signatory: Jens Hovgård Jensen