

EVALUATION CERTIFICATE

No. 0200-WL-07524 Revision 1

Object name	ECI
Object type	Non-automatic weighing indicator
Issued by	FORCE Certification EU - Notified Body No. 0200
In accordance with	EN 45501:2015, OIML R76:2006 and WELMEC Guide 8.8:2017 on metrological aspects of non-automatic weighing instruments.
Fractional factor (p_i)	0.5 (refer to 3.10.2.1 of EN 45501:2015).
Issued to	Esit Elektronik Sistemler İmalat ve Tic. Ltd. Şti. Nişantepe Mahallesi Gelin Çiçeği Sokak No: 36 Çekmeköy 34794 İstanbul TURKEY
Manufacturer	Esit Elektronik Sistemler İmalat ve Tic. Ltd. Şti.
In respect of	Non-automatic weighing indicators.
Characteristics	Suitable as a non-automatic weighing instruments with the following characteristics: Weighing range: Single-interval or multi-interval or multi-range. Accuracy class: III or IIII Verification scale interval: $e_i =$ Max_i/n_i Maximum number of verification scale intervals: $n_i =$ $\leq 10,000$ per interval/range Minimum input voltage per VSI: $1 \mu V$ The essential characteristics are described in the annex.
Description and documentation	The non-automatic weighing indicators are described and documented in the annex to this certificate.
Remarks	Summary of tests involved: See the annex to this certificate.

This evaluation certificate cannot be quoted in an EU type examination certificate without permission from the holder of the certificate mentioned above.

Note: This certificate is a revised edition which replaces previous revisions.

The annex comprises 12 pages.

Issued on 2020-11-11

Jens Hovgård Jensen
2020-11-11

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 120-33147.90.20 and ID no.: 0200-WL-09312

Signatory: Jens Hovgård Jensen