

# EU Type Examination Certificate

**No. 0200-NAWI-04001**

**IT10**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**        **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**        **Ets Torkhani Abdelkarim "ETA"**  
Rue de Mercure Z.I 2013,  
Ben Arous  
TUNISIA

**In respect of**    Non-automatic weighing instrument designated IT10 with variants of modules of load receptors, load cells and peripheral equipment.  
Accuracy class III and IIII  
Maximum capacity, Max: From 1 kg up to 300 000 kg  
Verification scale interval:  $e = \text{Max} / n$   
Maximum number of verification scale intervals:  $n \leq 6000$  for single-interval and multi-range (however, dependent on environment and the composition of the modules).  
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and conditions for certification are set out in the descriptive annex to this certificate.

The annex comprises 16 pages.

**Issued on**        **2018-03-22**  
**Valid until**     **2028-03-22**

  
**Signatory: J. Hovgård**