

# EU Type Examination Certificate

**No. 0200-NAWI-06147**

**FJ5 / GJW / HJW / HJP**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**        **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**        **Excell Precision Co., Ltd.**  
6F, No. 127, Lane 235, Pao-Chiao Rd.  
New Taipei City 23145  
TAIWAN

**In respect of**    A non-automatic weighing instrument designated the FJ5 / GJW / HJW / HJP with variants of modules of load receptors, load cells and peripheral equipment.  
Accuracy class III  
Maximum capacity, Max: 6 kg to 30 000 kg  
Verification scale interval:  $e_i = \text{Max}_i / n_i$   
Maximum number of verification scale intervals:  $n_i \leq 3750$   
(however, dependent on environment and the composition of the modules).  
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and examination conditions are set out in the descriptive annex to this certificate.

The annex comprises 13 pages.

**Issued on**        **2019-04-09**  
**Valid until**     **2029-04-09**

**Jens Hovgård Jensen**  
2019-04-09

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:

Task no.: 119-24481.90.20 and ID no.: 0200-NAWI-06147

**Signatory: J. Hovgård Jensen**