

# EU Type Examination Certificate

**No. 0200-NAWI-09749**

**Torrey L-EQ Series**

**NON-AUTOMATIC WEIGHING INSTRUMENT**

**Issued by**      **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

**Issued to**      **Fabricantes de Equipos para Refrigeración, S.A. de C.V.**  
Av. Día del Empresario 901, Col. Jardines de Guadalupe,  
C.P. 67115 Guadalupe, Nuevo León  
México

**In respect of**    Non-automatic weighing instrument designated Torrey L-EQ Series.  
Accuracy class III, single-interval  
Maximum capacity, Max: 4 kg to 20 kg  
Verification scale interval:  $e = \text{Max} / n$   
Maximum number of verification scale intervals:  $n \leq 4000$   
Variants of models are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and of OIML R76:2006.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 8 pages.

**Issued on**      **2021-02-12**  
**Valid until**    **2031-02-12**

**Jens Hovgård Jensen**  
2021-02-12

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:

Task no.: 120-35977.90.10 and ID no.: 0200-NAWI-09749

**Signatory: Jens Hovgård Jensen**