

# EU Type Examination Certificate

**No. 0200-MID-11226**

**FT-113 Fill**

**AUTOMATIC GRAVIMETRIC FILLING INSTRUMENT**

**Issued by**        **FORCE Certification**  
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/32/EU of the European Parliament and Council.

**Issued to**        **Flintec Transducers (Pvt) Ltd.**  
P.O: Box 24  
K.E.P.Z Phase I  
Spure Road 2  
Katunayake  
Sri Lanka

**In respect of**    An automatic gravimetric filling instrument designated **FT-113 Fill** with variants of modules of load receptors, load cells and peripheral equipment.  
Reference class 0.1  
Maximum capacity,  $Max = n \times e$   
Verification scale interval:  $e \geq 0.1 \text{ g}$   
Number of verification scale intervals:  $n \leq 10000$  per interval/range  
(however, dependent on environment and the composition of the modules)  
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in Annex 1 and the specific requirements in Annex VIII (MI-006), chapter I & III of the Directive 2014/32/EU is met by the application of OIML R61:2017 and WELMEC Guide 7.2:2018.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 11 pages.

**Issued on**        **2022-02-10**  
**Valid until**     **2032-02-10**

**Jens Hovgård Jensen**  
2022-02-10

Digitally signed by Jens Hovgård Jensen  
jhje@force.dk  
Certification Manager

FORCE Certification references:

Task no.: 121-30042.90.10 and ID no.: 0200-MID-11226-1

**Signatory: Jens Hovgård Jensen**