

EU Type Examination Certificate

No. 0200-NAWI-11227

FT-111 / FT-112 / FT-111 panel / FT-112 panel / FT-113

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements in Directive 2014/31/EU of the European Parliament and Council.

Issued to **Flintec Transducers (Pvt) Ltd.**
P.O: Box 24
K.E.P.Z Phase I
Spure Road 2
Katunayake
Sri Lanka

In respect of Non-automatic weighing instrument designated FT-111 / FT-112 / FT-111 panel / FT-112 panel / FT-113 with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class: III and IIII
Maximum capacity, Max: From 1 kg up to 500 000 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n_i \leq 10000$ for single-interval / multi-range / multi-interval (however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

The principal characteristics and conditions for certification are set out in the descriptive annex to this certificate.

The annex comprises 19 pages.

Issued on **2022-02-10**
Valid until **2032-02-10**

Jens Hovgård Jensen
2022-02-10

Digitally signed by Jens Hovgård Jensen
jhje@force.dk
Certification Manager

FORCE Certification references:

Task no.: 121-30042.90.20 and ID no.: 0200-NAWI-11227-1

Signatory: Jens Hovgård Jensen