

EVALUATION CERTIFICATE

No. 0200-WL-03388

Object name EM100-G
Object type Analog data processing device
Issued by Force Certification A/S

Issued in accordance with the requirements in WELMEC Guide 8.8:2017” Guide on General and Administrative Aspects of the Voluntary System of Modular Evaluation of Measuring instruments”.

In accordance with EN 45501:2015, OIML R76:2006 and WELMEC Guide 8.8:2017.

Issued to Flintec UK Ltd.
W4/5 Capital Point
Capital Business Park
Wentloog, Cardiff CF3 2PW
United Kingdom

Manufacturer Flintec, Katunayake, Sri Lanka

In respect of An analog data processing device tested as a module for a weighing instrument.

Characteristics Suitable for non-automatic weighing instruments with the following characteristics:
Weighing range: Single-interval or multi-range or multi-interval
Accuracy class: III or IIII
Number of VSIs: $n \leq 10,000$
Minimum input voltage per VSI: $\geq 0.3 \mu\text{V}$
The essential characteristics are described in the annex.


Description and documentation The weighing indicator is described and documented in the annex to this certificate.

Remarks Summary of tests involved: see annex.

This evaluation certificate cannot be quoted in an EU type examination certificate without permission of the holder of this certificate mentioned above.

The annex comprises 5 pages.

Issued on 2018-06-14


Signatory: J. Hovgård

FORCE Certification references:
Task no.: 117-29767 and ID no.: 0200-WL-03388